

波音特生物科技（南京）有限公司

日立 (HITACHI 7170)

LP(a)

		Analyze	Calib.	Range	STD Conc	
Assay/Point	2Point End		10	18	34	0
Wave(Sub/Main)			660			
S. Vol. (Normal)	3.0		0.0	0		
S. Vol. (Dec.)			0.0	0		
S. Vol. (Inc.)	6.0		0.0	0		
Diluent	Water		0			
Reagent Vol. (R1)	255		0	#	0	
Reagent Vol. (R2)	0		0	00000	0	
Reagent Vol. (R3)	85		0	#	0	
Reagent Vol. (R4)	0		0	00000	0	
ABS. Limit	32000		Inc.	Twin Test		
Prozone limit	0	0	Lower			
Cell Det.	Detergent 1					
		Analyze	Calib.	Range	STD Conc	
Calib. Type	SPLINE					
Point	5	Span Point			3	
Weight	0					
Auto Calibration						
	Timeout			Chg. Over		
Blank	0	Change Lot				
Span	0	Change Bottle				
2Point	0					
Full	0					
SD Limit	999					
Duplicate Limit	500					
Sensitiv. Limit	0					
S1ABS Limit	-32000		32000			
		Analyze	Calib.	Range	STD Conc	
<Standard>	(1)	(2)	(3)	(4)	(5)	(6)
Conc.	0.0	#	#	#	#	0
Position	#	#	#	#	#	0
Volume	3.0	3.0	3.0	3.0	3.0	0
<Pre-Diluent>						
Volume	0.0	0.0	0.0	0.0	0.0	0.0
Diluent	0	0	0	0	0	0
Calib. Code	0	0	0	0	0	0

“#” 由使用者自行输入浓度或位置。

南京市江宁区科学园兴民南路85号

电话：025-52425019 传真：025-52424836-807