

Bioway Chemistry Reagent Series

Complement C3 Test Kit


Detection of Complement C3 in Human Serum on Chemistry Analyzers



Cat. No. R014K11

Complement C3 Test Kit

SUMMARY OF TEST PROCEDURE



*Refer to Figure 1 and the package insert for detail

Table 1: Instrument Parameters*

Calibration method	6 point non-linear	Slope of reaction	increase
Wavelength	340 nm	Sample volume	3 µl
Test method	2 point end	R1 volume	240 µl
Reaction temperature	37°C	R2 volume	80 µl

INTENDED USE

Bioway Chemistry Reagent Series Complement C3 Reagent Kit (the Kit) is an immunoturbidimetric assay intended for *in vitro* quantitative detection of complement C3 in human serum on automated clinical chemistry analyzers.

SUMMARY AND EXPLANATION

Complement component 3 (C3) is the most abundant protein of the complement system. It plays a central role in the complement system and contributes to innate immunity. When activated by classical or alternative pathway, complement destroys foreign agents, release histamine, and release leukocytes from the bone marrow. C3 levels can be used in determining inherited or acquired deficiencies, including active lupus nephritis, severe infections, and inflammation. Increased levels are usually found after trauma, surgery, biliary obstruction, or focal glomerulosclerosis.

TEST PRINCIPLES

The Kit utilizes immunoturbidimetry to measure the C3 level in human serum. During the test, C3 in the sample binds with the specific anti-C3 antibody to cause agglutination. The turbidity caused by agglutination is detected optically by chemistry analyzer. The change in absorbance is proportional to the level of C3 in the sample. The actual concentration is obtained by comparing with a calibration curve with known concentrations.

MATERIALS PROVIDED

Reagents:

R1	Phosphate buffer, Polyethylene glycol, Sodium azide < 0.1%
R2	anti-C3 antibodies, Tris buffer, sodium azide < 0.1%

MATERIALS NEEDED BUT NOT PROVIDED

- Automated chemistry analyzer.
- C3 calibrator set (available for purchase) and control set (commercially available).

INSTRUMENT

The Kit is applicable on most automated chemistry analyzers. Refer to specific instrument application for suggested settings.

STORAGE AND STABILITY

Store the reagents at 2-8°C. Avoid direct sunlight. The Kit is stable through the expiration date when stored properly. R1 and R2 reagents are stable for 1 month at 2-8°C after opening.

PRECAUTIONS

- The Kit is for *in vitro* diagnostic use only. Not for use in humans or animals.
- The instructions must be followed to obtain accurate results.
- Do not use the reagents beyond the expiration date.
- Treat all specimens as infectious. Proper handling and disposal procedures of specimens and test materials should be strictly followed.
- Reagents contain less than 0.1% sodium azide as preservative; avoid contact with skin and eyes, flush with copious amounts of water when disposing.

SPECIMEN COLLECTION AND HANDLING

Follow standard laboratory procedures to collect serum samples. It is recommended to perform test immediately after sample collection. If the test cannot be done immediately, store sample at 2-4°C for up to 3 days or at -20°C for up to 1 months. Avoid repeated freezing and thawing.

TEST PROCEDURE (see Figure 1)

No pretreatment required for reagents and samples.

Calibration: 6 level calibrator set available for purchase. Recommend using Bioway calibrators for optimal results. Use multi-point non-linear calibration method.

Test procedure: see Figure 1 and Table 1 for instrument parameter setup. Refer to specific instrument application for suggested setting.

- Add 3 µl of sample and 240 µl of R1; mix well and incubate at 37°C for 300 seconds.
- Take optical density measurement OD 1 just before addition of R2.
- Add 80 µl of R2, mix well and incubate at 37°C.
- Take optical density measurement OD 2 at 600 seconds.
- Calculate $\Delta OD = OD 2 - OD 1$

RESULT

The C3 value can be obtained by using the calculated ΔOD to find the corresponding value on a calibration curve prepared with known values.

EXPECTED VALUES

80 - 185 mg/dL.

It is recommended for each laboratory to establish its own expected values.

QUALITY CONTROL

Using commercially available controls with known concentration is recommended before each batch of tests to ensure the test is properly

performed and all reagents and the instrument are functional as specified.

LIMITATIONS

1. The Kit is for *in vitro* use on automated chemistry analyzers only.
2. Hemolysis samples may cause inconsistent results.
3. The test result from the Kit should not be used as the only basis for definite diagnosis.

PERFORMANCE CHARACTERISTICS

Linearity: 0 – 335 mg/dL ($R \geq 0.995$)

Precision: Within Run: $CV \leq 6\%$;
Run-to-Run: $CV \leq 10\%$

Interference: no interference detected for: Bilirubin (60 mg/dL), triglycerides (1000 mg/dL), and hemoglobin (10 g/L).

Reagent Blank Absorbance: at 340 nm wavelength and 10 mm optical diameter, O.D. ≤ 0.10

REFERENCES

1. Burtis C., Ashwood, E.R.(ed). Tietz Textbook of Clinical Chemistry, 3rd ed. Philadelphia, PA; WB Saunders Co; 503-4; 1999.
2. De Bruijn MH, Fey GH (February 1985). "Human complement component C3: cDNA coding sequence and derived primary structure". *Proc. Natl. Acad. Sci. U.S.A.* 82 (3): 708–12.
3. Lachmann P (December 1975). "Genetics of the complement system". *J. Med. Genet.* 12 (4): 372–7.
4. Matsuyama W, Nakagawa M, Takashima H, Muranaga F, Sano Y, Osame M (December 2001). "Molecular analysis of hereditary deficiency of the third component of complement (C3) in two sisters". *Intern. Med.* 40 (12): 1254–8.
5. Sahu A, Lambris JD (April 2001). "Structure and biology of complement protein C3, a connecting link between innate and acquired immunity". *Immunol. Rev.* 180: 35–48.

Not Intended for Sale in the United States.

Pointe Biotech (Nanjing) Co., Ltd.

No.85,Xingmin South Road, Jiang Ning District, Nanjing,P.R.China 211100

Tel:86-25-52425019,

Fax:86-25-52424836

info@biowaydx.com